

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710258
----------------	--------------	--------------------	--------

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
35 M/ml	± 5.0 M/ml	242012351		2024/12/20	2026/12/20
Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*	
1 L - Bio	35.56	0.10	0.29	46.10-46.37	
1 L - QC	36.16	0.66	1.83	46.15-47.87	
5 ml - QC	36.47	0.16	0.43	47.21-47.62	
QC Analysis by: Carlos A Ulloa		Reviewed by: Nicole Thibodeaux			
Date: 2025/03/14		Date: 2025.04.25			
Signature: Carlos A Ulloa		Signature: Nicole Thibodeaux			

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.