Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710257

Concentration Type	Acceptable Range	Lot #	Date of Manufacture	Expiration Date
18 M/ml	± 2.5 M/ml	231310181	2023.10.13	2025.10.31

Test Type and Quantity	20 µm Fixed Coverslip Chamber Calibration	S.D.	CV%	Calculated Hemacytometer Calibration Range*
1 L - Bio	18.62	0.08	0.44	24.10-24.32
1 L - QC	17.68	0.48	2.68	22.37-23.60
5 ml - QC	18.73	0.19	0.99	24.10-24.59

QC Analysis by: Carlos A. Ulloa	Reviewed by: Nicole Mibockaw
Date: 24-06-17	Date: 2024.06.17
Signature: Carlos A. Ulloa	Signature: [signature]

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.