Innovations to Rely On
HAMILTON THORNE
info@hamiltonthorne.com
www.hamiltonthorne.com

CERTIFICATE OF ANALYSIS

Product	accu-beads®+	Part Number	710259

Concentration Type	Acceptable Range	Lot #		Date of Manufacture	Expiration Date
3 M/ml	± 1.0 M/ml	210207031		2021.02.07	2024.12.07
Test Type and Quantity	**20 µm Fixed Coverslip Chamber Calibration**	**S.D.**	**CV%**	**Calculated Hemacytometer Calibration Range***	
1 L - Bio	3.26	0.18	5.55	4.00-4.47	
1 L - QC	3.12	0.08	2.63	3.94-4.16	
5 ml - QC	3.28	0.19	5.65	4.03-4.51	

QC Analysis by:	Emily A. Wilson	Reviewed by:	Nicole Thibodeau
Date:	2022-12-07	Date:	2023.02.17
Signature:	Emily A. Wilson	Signature:	Nicole Th

* Calculation based on hemacytometer flow dynamics ratio as described in accu-beads Research Report 45. Please contact Hamilton Thorne for more information.